# NMR CE Y PLAN

This safety plan is to be considered an addendum to the Chemical Hygiene Plans of the Boston College Chemistry Department and Boston College. It does not supersede or contradict the Chemical Hygiene Plans of either. Hazards and procedures unique to the NMR Center are identified in this document, and precautions and prohibitions related to those hazards are specified. The safety plan is not to be considered exhaustive, and will be amended as additional hazards or better procedures are identified.

## 1. Electrical Hazards

The NMR spectrometers in the lab operate on either 240 volts AC or 120 volts AC and have several high-voltage DC components, all of which can be hazardous or fatal if accidental electrocution occurs. To prevent accidents, the following precautions and prohibitions shall apply to all NMR users and visitors to the NMR laboratory:

a. No person may access the instruments without proper training and authorization from the NMR Center director and Chemistry Department operations manager.
b. Extreme caution should be used whenever the instruments are being tuned or otherwise used in a way that makes it necessary to be near the console or magnet.
c. No user shall access the instrument panels or spectrometer consoles unless under the observation and guidance of the NMR Center director.
d. Any accumulation of water around or near the instruments should be reported to the NMR Center director and the wet areas should be avoided to prevent electrocution.

d

cryogens, and to avoid asphyxiation in the event of a magnet quench, the following precautions and prohibitions shall apply to all NMR users and visitors to the NMR laboratory:

a.

E

## 5. Chemical Hazards

The following precautions and prohibitions apply to all NMR users and visitors to the NMR laboratory:

a. NMR Solvents must be handled as specified in the Chemical Hygiene Plan of Boston College. Because there are no fume hoods in the NMR laboratory, samples requiring a hood for safe handling must be prepared outside of the NMR laboratory.
b. Chemical spills or accidental exposure to NMR solvents must be reported to the NMR Center director and Chemistry Department operations manager.
c. Material Data Safety Sheets are available for all standard and calibration samples used in the NMR laboratory.

## 6.